



**CITY OF ST. LOUIS
DEPARTMENT OF FINANCE
OFFICE OF THE SUPPLY COMMISSIONER**

**CAROL L SHEPARD, CPA
SUPPLY COMMISSIONER**

**FRANCIS G. SLAY
MAYOR**

**1200 MARKET ST RM 324
ST LOUIS MO 63103
PHONE: 314-622-4580
FAX: 314-622-4141**

ADDENDUM NO. 1

33516Q0003 – Autopsy Saw
Bid Opening Date – December 15, 2015

To: All Bidders
From: Dotlyn Bedessie, Buyer
Date: December 8, 2015
Re: 33516Q0003 – Autopsy Saw

This addendum shall be considered part of the Specifications for bid (except as noted otherwise) and is issued to change, amplify, or delete from or otherwise explain these documents where provisions of this addendum differ from those of the original documents. This addendum shall have precedence over the original documents and shall govern.

Q1: The Mortech saw is no longer manufactured and we do not have one (or aware of one) that is immersible in water.

AI: The specification requesting a submersible saw is no longer valid. Any Autopsy Saw like the Stryker, Salam's Model 1200, or acceptable equal is adequate.

All bidders must attach this form to the original Bid/Quote Form. This completed form will serve as acknowledgement that this information has been received.

ADDENDUM RECEIVED AND ACKNOWLEDGE

Name of Firm

Representative

Signature

Date